

Pairing the Philips Thermometer with MyLife365 Smart Devices

Follow Manufacturer Instructions for Use.

Still have questions or need assistance? Contact Support at support@life365.health

Pairing the Philips Thermometer with MyLife365 Smart Devices

After assigning a smart device to a member, select the Reminders application. You will receive two prompts, select "ALLOW" for both prompts.

This step must be completed before pairing Bluetooth Devices.






Select the Bluetooth Icon*.

Enter the password "3651" and select "OK".


*the Bluetooth App Icon location may vary by device






On the "My Devices" page, select "Add New" (A).

The smart device will search for the Bluetooth device.





Pairing the Philips Thermometer with MyLife365 Smart Devices





Press the power button to turn **ON** the thermometer.


Allow a few moments for the Bluetooth pairing application to find the thermometer.

Once found, select the Bluetooth device **name** (B).

Select "Add to My Devices" (C).

This is an example. Device name displayed will reflect the Bluetooth device the user is attempting to pair.

Device pairing is now complete.




MANUAL PAIRING:

Difficulty pairing this device?

- Try manually pairing the device by selecting "Add New" on the "My Devices" page (A), then selecting "Manually" (B).
- Search for the device by Manufacturer by selecting the magnifying glass (C).
- Select the device, then select "Add to My Devices".

Device pairing is now complete.

Take a test reading with the device. Hub will announce the reading.



NOTE:

The Reminders Application must be opened BEFORE opening the Bluetooth Application. If the Bluetooth Application is opened before the Reminders Application, a "Choose Server Location" error screen will display.

If this error screen is shown:

- 1. Exit the Bluetooth Application
- 2. Enter the Reminders Application
- 3. Accept both prompts
- 4. Re-enter the Bluetooth Application
- 5. Select the Refresh symbol (D).

You can now complete device pairing.

